Brillenmann®

Manufacturer

Brillenmann GmbH & Co. KG

Mollenbachstraße 6

71229 Leonberg

GERMANY

Single-Registration-Number (SRN)

DE-MF-000023069

We declare under sole responsibility that the products

Produktname

Spectacle Frames / Models see appendix

Marken: K16 / Sim-Sala-Bim / PASS / +choc- / XXL / Xclusiv /Onda

Basic UDI-DI

4099911frames_mix_fullVX

Nomenclature

GMDN Code 32816

EMDN Code Q02100203

Risk Class

Class I – non sterile, no meassuring functions

Conformaty assessment procedure

Pursuant to article 52 (7) for medical devices

Conforms with the following regulations:

Regulation (EU) 2017/745 of the European Parliament and of the Council on medical devices.

Regulation (EC) No 1907/2006 on the Registration, Evaluation, Authorization and Restriction of Chemicals (REACH)

Complies with European standard EN ISO 12870

We operate a systematic procedure to monitor the product after placing it on the market.

Date and Location

Ana Teixeira

MDR Responsible Person

Joao Teixeira